

CERTIFICATE OF CONFORMANCE

AIR ZERO GRADE G Product Code: 384150

This is to certify that quality verification tests have been performed on representative samples of Zero Air from Coregas' approved production plants and the results of the tests comply with the requirements of the Zero Air specification.


Component	Chemical Formula	Specification	Unit
Nitrogen	N ₂	Balance	
Oxygen	O ₂	21.0 ± 0.5	%
Moisture	H ₂ O	≤ 15	ppm
Hydrocarbons as methane	CH ₄	≤ 1	ppm

Cylinder connection as per AS2473
Cylinder pressure at 15°C
Package content at 15°C, 101 kPa

Type 60
20,000 kPa
9.4 m³

Remarks: -

Approved by:



National Quality and Compliance Manager

Date of issue: 1 September 2018



Complies with requirements
of ISO 9001

ISO 9001

Coregas Pty Ltd

66 Loftus Road, Yennora NSW 2161, Australia
Phone (+61) 1800 807 203 | coregas.com.au